

AD-A197 165

OFFICE OF NAVAL RESEARCH

Contract N00014-86-K-0639

R&T Code 4133013...2

Technical Report No. 4

Low Frequency Laser Debye-Sears
Ultrasonic Measurements

by

G.R. Phillips, C.N. Merrow, J.K. Mitchell,
R.M. Barton, and E.M. Eyring^{*}

Prepared for Publication
in the
Review of Scientific Instruments

University of Utah
Department of Chemistry
Salt Lake City, UT 84112

June 30, 1988

DTIC
ELECTE
JUL 14 1988
S D
ave
E

Reproduction in whole or in part is permitted for
any purpose of the United States Government

This document has been approved for public release
and sale; its distribution is unlimited.

ADA197165

REPORT DOCUMENTATION PAGE

1a. REPORT SECURITY CLASSIFICATION Unclassified			1b. RESTRICTIVE MARKINGS		
2a. SECURITY CLASSIFICATION AUTHORITY			3. DISTRIBUTION / AVAILABILITY OF REPORT Approved for public release and sale. Distribution unlimited.		
2b. DECLASSIFICATION / DOWNGRADING SCHEDULE			5. MONITORING ORGANIZATION REPORT NUMBER(S)		
4. PERFORMING ORGANIZATION REPORT NUMBER(S) ONR Technical Report No. 4			7a. NAME OF MONITORING ORGANIZATION Office of Naval Research Resident Representative		
6a. NAME OF PERFORMING ORGANIZATION University of Utah		6b. OFFICE SYMBOL (If applicable)		7b. ADDRESS (City, State, and ZIP Code) University of New Mexico Bandelier Hall West Albuquerque, NM 87131	
6c. ADDRESS (City, State, and ZIP Code) Department of Chemistry University of Utah Salt Lake City, UT 84112		8b. OFFICE SYMBOL (If applicable) ONR		9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER N00014-86-K-0639	
8a. NAME OF FUNDING / SPONSORING ORGANIZATION Office of Naval Research		8c. ADDRESS (City, State, and ZIP Code) 800 N. Quincy Street Arlington, VA 22217		10. SOURCE OF FUNDING NUMBERS	
		PROGRAM ELEMENT NO.		PROJECT NO.	TASK NO.
					WORK UNIT ACCESSION NO.
11. TITLE (Include Security Classification) (U) Low Frequency Laser Debye-Sears Ultrasonic Measurements					
12. PERSONAL AUTHOR(S) G.R. Phillips, C.N. Merrow, J.K. Mitchell, R.M. Barton, and E.M. Eyring*					
13a. TYPE OF REPORT Technical		13b. TIME COVERED FROM 9/87 TO 6/88		14. DATE OF REPORT (Year, Month, Day) 1988, June 30	
15. PAGE COUNT 10					
16. SUPPLEMENTARY NOTATION Prepared for publication in Review of Scientific Instruments					
17. COSATI CODES			18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)		
FIELD	GROUP	SUB-GROUP	laser Debye-Sears technique, ultrasonic absorption, quartz piezoelectric transducers, crown ethers.		
19. ABSTRACT (Continue on reverse if necessary and identify by block number)					
<p>The frequency range accessible to the laser Debye-Sears technique has been extended downward to 7 MHz. Modifications include a 1 MHz piezoelectric transducer, a 5 mW HeNe probe laser, and an extended optical path. The precision is experimentally characterized using the complexation of NaCl and 18-crown-6 in H₂O. The accessibility of lower frequencies permits the detection of a previously unobserved second relaxation.</p>					
20. DISTRIBUTION / AVAILABILITY OF ABSTRACT <input checked="" type="checkbox"/> UNCLASSIFIED/UNLIMITED <input type="checkbox"/> SAME AS RPT <input type="checkbox"/> DTIC USERS			21. ABSTRACT SECURITY CLASSIFICATION Unclassified		
22a. NAME OF RESPONSIBLE INDIVIDUAL			22b. TELEPHONE (Include Area Code)		22c. OFFICE SYMBOL

Low Frequency Laser Debye-Sears Ultrasonic Measurements

Gregory R. Phillips, Clifton N. Merrow,^{a)} Jerry K. Mitchell,³

R. Michael Barton, and Edward M. Eyring*

Department of Chemistry

University of Utah

Salt Lake City, Utah 84112

on For	
RA&I	<input checked="" type="checkbox"/>
nced	<input type="checkbox"/>
ation	
By _____	
Distribution/	
Availability Codes	
Dist	Avail and/or Special
A-1	

ABSTRACT

The frequency range accessible to the laser Debye-Sears technique has been extended downward to 7 MHz. Modifications include a 1 MHz piezoelectric transducer, a 5 mW HeNe probe laser, and an extended optical path. The precision is experimentally characterized using the complexation of NaCl and 18-Crown-6 in H₂O. The accessibility of lower frequencies permits the detection of a previously unobserved second relaxation.



Received by DTIC on 10/1/78
Clifton N. Merrow, Jr.
Edward M. Eyring

INTRODUCTION

Ultrasonic absorption techniques cover a wide time range (10^{-5} - 10^{-10} seconds). The difficult and time consuming nature of many ultrasonic methods, and the need for more than one ultrasonic instrument to cover the desired 0.5 MHz to 500 MHz frequency range, has limited the use of ultrasonic techniques. Of the several ultrasonic methods, the laser Debye-Sears technique is perhaps the easiest and quickest to use. Eyring and coworkers¹ developed an automated system capable of measuring frequencies between 15 and 300 MHz. Automation results in shorter measurement time and higher precision than is typically obtained with pulsed systems over the same frequency range.²

Sound waves produce periodic variations in the density of a condensed phase which can act as a three dimensional diffraction grating for a light beam. The strength of the acoustic signal can be monitored using a laser beam diffracted by this grating. The acoustic absorption coefficient, α , is calculated from the intensity of the diffracted beam as a function of the distance between the transducer and the laser beam. The ultrasonic frequency dependence of α contains kinetic and thermodynamic information about chemical reactions occurring in the sample medium. Stuehr² discusses this relationship in detail.

At Utah, a 5 MHz transducer has been excited at its odd harmonics to determine α at frequencies above 15 MHz. When necessary, ultrasonic data between 1 and 15 MHz have been obtained using the resonance technique in

collaboration with other researchers.³ This note reports an extension of the laser Debye-Sears technique to frequencies below 15 MHz. This is accomplished with a 1 MHz transducer using a HeNe laser rather than the more expensive Ar^+ laser used in previous work. Reexamination of the complexation of NaCl with 18-crown-6 in aqueous solution over a lower frequency range reveals a second relaxation which was previously undetected. The lower frequency range has also proven useful in similar rate studies of reactions in dry aprotic solvents.

I. ULTRASONIC APPARATUS

A schematic of the optical path used to separate the diffracted and undeflected beams is shown in Fig. 1. The ultrasonic path length, the angle of the laser beam relative to the transducer, and the positioning of the moveable mirror are varied using stepping motors (Superior Electric) controlled from an IBM XT-286 computer with routines written in the C language. The optical path length is 16 feet; the entire apparatus occupies 12 feet of bench top.

The transducer is an X-cut quartz crystal (Valpey-Fisher) with a fundamental frequency of 1 MHz and a diameter of one inch. The output of a Hewlett-Packard 8656B synthesized signal generator is amplitude modulated before being applied to the transducer. A PARC 124 lock-in detector measures the signal from the photomultiplier.

A 5 mW HeNe laser (CW Radiation) monitored the acoustic strength, in contrast to a 200 mW Ar^+ laser used previously.¹ Even with the lower laser

power, the diffracted beam required attenuation before detection. (For example, at 9 MHz the beam was attenuated by a factor of 20). With this arrangement, α has been measured at frequencies up to 71 MHz.

II. REAGENTS

Sodium chloride (Mallinckrodt) was dried at 70°C before use. 18-Crown-6 (Parish) was purified using standard methods.⁴ Solutions were prepared with deionized water.

III. RESULTS AND DISCUSSION

The precision of low frequency results has been evaluated using an aqueous solution of 0.30 M NaCl and 0.096 M 18-crown-6 at 25°C. The intensity of the diffracted light as a function of the ultrasonic path length is

$$I(x) = I(0) \exp(-2\alpha x) \quad (1)$$

Here x is the ultrasonic path length in centimeters, α is the acoustic absorption coefficient (in cm^{-1}), and $I(x)$ is the intensity of the diffracted beam. Estimates of α , denoted $\hat{\alpha}$, and its standard deviation, denoted $s(\hat{\alpha})$, are calculated from the slope of a linear regression of the logarithm of $I(x)$ versus x . Measurement errors in the light intensity and transducer position, as well as deviations from a straight line relationship are included in $s(\hat{\alpha})$. Consequently, the relative standard deviation, $s(\hat{\alpha})/\hat{\alpha}$, reflects the overall precision in $\hat{\alpha}$. The average percentage

relative standard deviation at several frequencies is shown in Table I. Also listed are the mean values of α at these frequencies and the mean deviation between replicate measurements as a percentage of the mean. For a small number of observations, the mean deviation performs better than the root mean square deviation. Both measures of precision in Table I demonstrate the reliability of this method at low frequencies.

The complexation of NaCl with 18-crown-6 in aqueous solution has previously been studied⁵ using the Debye-Sears technique at frequencies above 15 MHz. Results on a solution containing 0.30 M NaCl and 0.096 M 18-crown-6 revealed a single relaxation at 15.2 MHz. In the absence of a resonator cell, examination of lower frequencies was not then possible. Additional laser Debye-Sears data have now been obtained under identical conditions using a 1 MHz quartz transducer. Figure 2 presents the combined laser Debye-Sears data from a 5 MHz crystal and a 1 MHz crystal. Analysis of the data in Fig. 2 using a nonlinear least squares algorithm reveals the presence of two relaxations at 11.2 MHz and 28 MHz. The data have been plotted as excess absorption to emphasize low frequencies. Similar conclusions are reached from the analysis of Debye curves.²

Eliminating the need for an ultrasonic resonator for all but the very lowest frequencies is not the only potential benefit of this work. The Debye-Sears method gives a direct measure of α . The resonator method, on the other hand, requires correction terms for mechanical loss, and can produce systematic errors.^{6,7} The Debye-Sears method is not suitable for solutions of low absorption, such as pure water. Such cases are easily

recognized by a large value of $s(\hat{\alpha})$, and can be avoided. However, this is not usually a problem in the study of kinetic processes.

ACKNOWLEDGMENTS

The authors thank D. P. Cobranchi for providing a sample of purified 18-crown-6. This research was supported in part by the Office of Naval Research.

FOOTNOTES AND REFERENCES

a) Currently an NRC Postdoctoral Research Associate. Ballistic Research Laboratory, Aberdeen Proving Ground, Maryland 21005-5066.

1. M.M. Farrow, S.L. Olsen, N. Purdie, and E.M. Eyring, *Rev. Sci. Instrum.* **47**, 657 (1976).
2. J. Stuehr in *Techniques of Chemistry*, 3rd ed., edited by G.G. Hammes, (Wiley-Interscience, New York, 1974), Vol. VI, Part II, Chap. 7.
3. L.J. Rodriguez, G.W. Liesegang, M.M. Farrow, N. Purdie, and E.M. Eyring, *J. Phys. Chem.* **82**, 647 (1978).
4. G.W. Gokel and D.J. Cram, *J. Org. Chem.* **39**, 2445 (1974).
5. G.W. Liesegang, M.M. Farrow, F.A. Vazquez, N. Purdie, and E.M. Eyring, *J. Am. Chem. Soc.* **99**, 3240 (1977).
6. Y. Yoshio and Y. Harada, *Jpn. J. Appl. Phys.* **19**, 1563 (1980).
7. F. Eggers and T. Funck, *Rev. Sci. Instrum.* **44**, 969 (1973).

Table I. Precision of α versus Frequency

f , MHz	$\bar{\alpha}$, $\text{cm}^{-1\text{a}}$	RSD, $^{\text{b}}$ %	RMD, $^{\text{c}}$ %
7.	0.178	3.8	2.6
9.	0.253	0.7	3.0
11.	0.344	0.8	0.3
13.	0.430	0.8	1.2
15.	0.489	0.8	0.9
17.	0.520	0.9	2.2
19.	0.571	0.6	0.8
21.	0.646	0.7	0.7
23.	0.715	1.0	0.4

^a Average of 3 - 5 replicate measurements at each frequency.

^b Relative Standard Deviation. $\text{RSD} = (100\%/n) \cdot \sum_i s(\hat{\alpha}_i)/\hat{\alpha}_i$, where the summation is over n replicate measurements.

^c Relative Mean Deviation. $\text{RMD} = (100\%/n) \cdot \sum_i \{|\alpha_i - \bar{\alpha}|/\bar{\alpha}\}$.

Figure Captions

Figure 1. Schematic diagram of the optical path used to separate the diffracted and undeflected beams. Neutral density filters are symbolized by ND. For clarity, no connections with the XT computer are shown.

Figure 2. Plot of $\alpha\lambda$ as a function of ultrasonic frequency, where λ is the wavelength of sound in solution. Measurements using a 1 MHz transducer(o) are consistent with those using a 5 MHz transducer(Δ) from reference 5. The solid line is the sum of two relaxations.



